

# Notice of Allowability

Application No.

10/086,753

Examiner

(Jackie) Tan-Uyen T. Ho

Applicant(s)

KUMAR, RAVI

Art Unit

3731

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to 2/16/2007.
2. ☒ The allowed claim(s) is/are 1-3,5-8,10,12,15-18 and 20-27.
3. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
  - a) ☐ All b) ☐ Some\* c) ☐ None of the:
    1. ☐ Certified copies of the priority documents have been received.
    2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
    3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

\* Certified copies not received: \_\_\_\_\_.


Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.

**THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.**

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
  5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
    - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
      - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date \_\_\_\_\_.
    - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date \_\_\_\_\_.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

## Attachment(s)

1. ☒ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☐ Information Disclosure Statements (PTO/SB/08), Paper No./Mail Date \_\_\_\_\_
4. ☐ Examiner's Comment Regarding Requirement for Deposit of Biological Material
5. ☐ Notice of Informal Patent Application
6. ☐ Interview Summary (PTO-413), Paper No./Mail Date \_\_\_\_\_
7. ☒ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other \_\_\_\_\_

  
(Jackie) Tan-Uyen T. Ho  
Primary Examiner  
Art Unit: 3731

## EXAMINER'S AMENDMENT

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

The application has been amended as follows:

Claim 23, after "recited in" replace "claim 23" with --claim 22--.

Claim 25, after "according to" replace "claim 25" with --claim 24--.

Claim 26, after "according to" replace "claim 26" with --claim 25--.

Claim 27, after "according to" replace "claim 25" with --claim 24--.

2. The following is an examiner's statement of reasons for allowance: Claims 1-3, 5-8, 10, 12, 15-18, 20-27 are allowed because the prior art fails to disclose or suggest an apparatus for occluding a blood vessel having a severed end and a lumen extending therefrom with a longitudinal axis wherein the apparatus comprises in combination of an insertion and a plug as claimed. Specifically, a plug having plurality of spokes as claimed and a tapered outer surface and a large diameter section of the plug together define a continuous occluding surface that occludes blood flow through the lumen of the blood vessel.

Purdy and Mebin-Uddin fail to disclose a plug with a continuous occluding surface. In order to incorporate an insertion device in view of Mebin-Uddin, the Purdy's plug would have a through hole thus there would not be a continuous occluding surface.

All references cited in the PTO-892 attached herewith fail to disclose or suggest a plug including a plurality of spokes.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to (Jackie) Tan-Uyen T. Ho whose telephone number is 571-272-4696. The examiner can normally be reached on MULTIFLEX Mon. to Sat..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, ANHTUAN NGUYEN can be reached on 571-272-4963. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



(Jackie) Tan-Uyen T. Ho  
Primary Examiner  
Art Unit 3731

May 1, 2007